



Unit 4C | Woodhall Business Park | Sudbury | Suffolk | CO10 1WH | United Kingdom
T: +44 (0) 1787 882038 | F: +44 (0) 1787 882039 | E: info@modusgauges.com | W: modusgauges.com

STANDARD TEST CERTIFICATE NO : 30558

TO :

INSTRUMENT SERIAL NO : C16124-1

OUR JOB NO : 16124

ITEM	QTY	INSTRUMENT DESCRIPTION WITH IDENTIFICATION/REF/PART CODE
1	1	50MM DIAMETER COMPOUND GAUGE STAINLESS STEEL CASE BRASS INTERNALS/WETTED PARTS DIRECT BACK ENTRY 1/4 BSP THREAD -30"HG/+30PSI & BAR GLYCERINE FILLED CLASS 2.5 PRODUCT CODE B122E103K

We certify that the above instrument has been tested to BS EN 837-1: 1998. Accuracy tested to and within \pm 2.50% FSD limits.

Instrument tested with: WATER

Instrument tested in: VERTICAL position

Laboratory Temperature 20°C \pm 2°C

CALIBRATION EQUIPMENT USED

0/20 BAR TRANSDUCER SERIAL NO: 34181

-1/0 BAR+"HG MASTER TEST GAUGE SERIAL NO: 312334

Calibration equipment above is within its period of calibration and traceable to NPL/National Standards

Signed:

Date: