



Unit 4C | Woodhall Business Park | Sudbury | Suffolk | CO10 1WH | United Kingdom
T: +44 (0) 1787 882038 | F: +44 (0) 1787 882039 | E: info@modusgauges.com | W: modusgauges.com

POINT TO POINT CALIBRATION CERTIFICATE NO: 27763

TO :

INSTRUMENT SERIAL NO : C14677-3

OUR JOB NO : 14677

ITEM	QTY	INSTRUMENT DESCRIPTION WITH IDENTIFICATION/REF/PART CODE
1	1	100MM DIAMETER PRESSURE GAUGE, BLACK STEEL CASE, BRASS INTERNALS/WETTED PARTS, DIRECT BOTTOM ENTRY, 3/8 BSP THREAD, 0-11 BAR/PSI, CLASS 1.0, DRY, PRODUCT CODE B401A348L

INSTRUMENT READING SCALED BAR	TEST PRESSURE READING RISING BAR	TEST PRESSURE READING FALLING BAR
0	0	0
2	2.021	2.001
4	4.007	3.982
6	6.025	6.004
8	8.050	8.022
10	10.046	10.048
11	11.060	-

We certify that the above instrument has been tested to BS EN 837-1: 1998. Accuracy tested to and within $\pm 1.00\%$ FSD limits. Readings are as shown above.

Instrument tested with: WATER

Instrument tested in: VERTICAL position

Laboratory Temperature 20°C $\pm 2^\circ\text{C}$

Calibration Equipment Used:

Transmitter:	0-20 Bar
Serial No:	34181
Accuracy:	0.05%

Calibration equipment above is within its period of calibration and traceable to NPL/National Standards

Signed:

Date: