



Unit 4C | Woodhall Business Park | Sudbury | Suffolk | CO10 1WH | United Kingdom
T: +44 (0) 1787 882038 | F: +44 (0) 1787 882039 | E: info@modusgauges.com | W: modusgauges.com

STANDARD TEST CERTIFICATE NO : 27752

TO :

INSTRUMENT SERIAL NO : C14673-1

OUR JOB NO : 14673

ITEM	QTY	INSTRUMENT DESCRIPTION
1	1	100MM DIAMETER PRESSURE GAUGE STAINLESS STEEL CASE BRASS INTERNALS/WETTED PARTS DIRECT BOTTOM ENTRY 1/2" BSP THREAD 0-15 PSI/BAR CLASS 1.6 GLYCERINE FILLED PRODUCT CODE: B422A240M

We certify that the above instrument has been tested to BS EN 837-1: 1998. Accuracy tested to and within $\pm 1.60\%$ FSD limits.

Instrument tested with: AIR

Instrument tested in: VERTICAL position

Laboratory Temperature $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$

Calibration Equipment Used:

Digital Pressure Gauge:	0-2.5 Bar
Serial No:	110811304D
Accuracy:	0.05%

Calibration equipment above is within its period of calibration and traceable to NPL/National Standards

Signed:

Date: